



testo 330 -1, -2, -3 / testo 330 -1 **L**, -2 **L**  
Flue gas analyser

Instruction manual

en



## K.3 Other device data

Characteristic	Values	
Operating temperature range	-5...45 °C	
Storage / transport temperature	-20...50 °C	
Ambient humidity	0...90% RH, not condensing	
Power supply	Battery block: 3.7 V / 2.4 Ah Mains unit: 6.3 V / 1.2 A	
Dimensions (L x W x H)	270 x 90 x 65 mm	
Weight	600 g (excluding battery)	
Memory	testo 330-1, -2 / testo 330-1 LL: 200 locations testo 330-3 / testo 330-2 LL: 400 locations	
Display	Monochrome, 4 grey levels, 160 x 240 pixels	
Battery storage temperature:	±0...35 °C	
Battery life	> 6h (pump on, display lighting off, 20 °C ambient temperature)	
Battery charge time	approx. 5-6 h	
Warranty	testo 330-1, -2, -3: Measuring instrument: 24 months Measuring cells: 24 months  Flue gas probe: 24 months Thermocouple: 12 months Battery: 12 months Warranty conditions: see <a href="http://www.testo.com/warranty">www.testo.com/warranty</a>  testo 330-1 LL, -2 LL: Measuring instrument: 48 Monate LL-Measuring cells O <sub>2</sub> , CO: 48 Monate, Other Measuring cells: 24 Monate Flue gas probe: 48 Monate Thermocouple: 12 Monate Battery: 12 Monate	
Option Bluetooth®	Typ-designation: Bluetooth Qualified Product Notice: Bluetooth listing identifier: Bluetooth listing company:	BlueNiceCom IV BNC4_HW2x_SW2xx B013784 10274
Option Bluetooth®	Range <10m	
Option Bluetooth®	EU countries	
Certification	Belgium (BE), Bulgaria (BG), Denmark (DK), Germany (DE), Estonia (EE), Finland (FI), France (FR), Greece (GR), Ireland (IE), Italy (IT), Latvia (LV), Lithuania (LT), Luxembourg (LU), Malta (MT), Netherlands (NL), Austria (AT), Poland (PL), Portugal (PT), Romania (RO), Sweden (SE), Slovakia (SK), Slovenia (SI), Spain (ES), Czech Republic (CZ), Hungary (HU), United Kingdom (GB) and Republic of Cyprus (CY). Other EFTA Countries Iceland, Liechtenstein, Norway and Switzerland Non-European countries Colombia, Turkey, El Salvador Information from FCC (Federal Communications Commission) - This instrument fulfils 15 of the FCC guideline - Commissioning is subject to the following two conditions: 1. This instrument may not cause any dangerous interference 2. This instrument must accept any interference received, including interference which may cause undesired operation. Changes The FCC requires that the user is informed that all changes and modifications to the instrument which are not expressly approved by Testo AG, can void the right of the user to operate the instrument.	