

Selection and ordering data

Description	Quantity for 2 years	Quantity for 5 years	Article No.
Analyzer unit			
Reference gas supply (pump, restrictor, pressure switch, hose)	1	1	A5E00114838
O-ring	1	2	C74121-Z100-A6
Pressure switch (sample gas)	1	2	C79302-Z1210-A2
Flowmeter	1	2	C79402-Z560-T1
Sample chamber			
• Stainless steel, mat. no. 1.4571; non-flow-type compensation branch	-	1	C79451-A3277-B535
• O-ring (measuring head)	2	4	C79121-Z100-A32
• O ring (fitting)	2	4	C71121-Z100-A159
Measuring head (non-flow-type compensation branch)	1	1	C79451-A3460-B525
Restrictor for sample gas path, hose	2	2	C79451-A3480-C10
Reference gas path, 3000 hPa (set of parts)	1	1	C79451-A3480-D518
Electronics			
Front plate with keyboard	1	1	A5E00259978
Motherboard, with firmware: see spare parts list	-	1	
Adapter plate, LCD/keyboard	1	1	C79451-A3474-B605
Magnetic field connection plate	-	1	C79451-A3474-B606
LC display	1	1	W75025-B5001-B1
Connector filter	-	1	W75041-E5602-K2
Fuse			
• 0.63 A/250 V (230 V version)	2	3	W79054-L1010-T630
• 1.0 A/250 V (110 V version)	2	3	W79054-L1011-T100

If the OXYMAT 61 was supplied with a specially cleaned gas path for high oxygen context ("Clean for O₂ service"), please ensure that you specify this when ordering spare parts. This is the only way to guarantee that the gas path will continue to comply with the special requirements for this version.